#### **PRIMARY**



#### Implant Summary Report for: ConforMIS

#### iTotal G2 XE and iTotal G2 (Bicondylar tray)

Comprising PRIMARY knees implanted up to: 07 December 2017

NJR Database extract: 05 February 2018

Produced on: 10 February 2018

Licenced for use until: 10 February 2019

#### **Contents**

Recorded Usage in NJR	2
Patient and Procedure Details	3
Revision and Survivorship	5
APPENDIX A – Component List	13

This report has been produced by Northgate Public Services (UK) Ltd. It summarises usage and outcomes associated with the iTotal G2 XE and iTotal G2 (Bicondylar tray), based on data collected by the NJR up to the specified date for the components listed in Appendix A.

# **NJR Recorded Usage**

#### **Implant Usage**


Date of first recorded usage in the NJR: 09 November 2011


Date of last recorded usage in this dataset: 07 December 2017

Maximum implantation time: 5.6 years

Mean implantation time: 1.9 years

Totals Recorded in NJR	Cumulative Total
Procedures	576
Patients	504
Centres	46
Consultants	45
Implanting Surgeons	47





	Year of implantation														
Current Outcome	Pre- 2006	2006	2007	2008	2009	2010	2011	2012	2013	2014	2015	2016	2017	Total	%
Death	0	0	0	0	0	0	0	1	1	0	1	1	0	4	0.7%
Revised	0	0	0	0	0	0	1	0	0	2	0	0	0	3	0.5%
Unrevised	0	0	0	0	0	0	0	15	26	76	124	171	157	569	98.8%
Total	0	0	0	0	0	0	1	16	27	78	125	172	157	576	100%



# **Patient / Procedure Details**

## **Patient Details**

	iTotal G2 XE and iTotal G2 (Bicondylar tray)	All TKR in NJR
Total Procedures	576	958,398
Total Patients	504	786,185
Demographics		
Mean age	66.6	70.1
< 50	5.0%	1.8%
50 – 59	16.7%	11.2%
60 – 69	38.2%	32.4%
70 – 79	32.5%	38.5%
≥ 80	7.6%	16.0%
Median BMI	29	30
% BMI information available	78.6%	60.2%
Underweight ( BMI < 18.5)	0.2%	0.3%
Normal (18.5 ≤ BMI < 25)	14.6%	10.2%
Overweight (25 ≤ BMI < 30)	38.4%	34.2%
Obese I (30 ≤ BMI < 35)	30.2%	32.3%
Obese II (35 ≤ BMI < 40)	10.4%	16.2%
Obese III (BMI ≥ 40)	6.2%	6.9%
% Male	51.6%	42.6%
ASA Grades		
P1 - Fit and healthy	21.4%	11.0%
P2 - Mild disease not incapacitating	68.1%	72.4%
P3 - Incapacitating systemic disease	10.6%	16.2%
P4 / P5	0.0%	0.3%
Indications		
Osteoarthritis	98.61%	97.32%
Rheumatoid Arthritis	0.52%	1.63%
Other Inflammatory Arthropathy	0.52%	0.65%
Previous Trauma	0.35%	0.56%
Avascular Necrosis	0.17%	0.33%
Other	0.52%	0.36%

Total of indications may exceed total number of implants, since more than one indication can be listed per case.




# **Patient / Procedure Details**

# **Intraoperative Adverse Events**

		Event F		
Adverse Event	Number <sup>†</sup>	iTotal G2 XE and iTotal G2 (Bicondylar tray)	All other TKR in NJR	p value
None	575	99.83%	99.41%	0.275
Fracture	0	0.00%	0.16%	1
Patella Tendon Avulsion	1	0.17%	0.04%	0.226
Ligament Injury	0	0.00%	0.07%	1
Other	0	0.00%	0.34%	0.275


<sup>†</sup> multiple events may be listed for one procedure

<sup>\*</sup> As percentage of procedures for which adverse event data was recorded





#### **Patient - Time Incidence Rate**



	PTIR (95% Confidence Interval)
iTotal G2 XE and iTotal G2 (Bicondylar tray)	0.280 (0.058 - 0.817)
All TKR in NJR	0.392
Status	Not an outlier


Outlier level 1: Used in ≥ 100 primary operations and PTIR 95% lower confidence > 2 x group PTIR

Outlier level 2: Used in ≥ 100 primary operations and PTIR 95% lower confidence > 1.5 x group PTIR OR

Used in < 100 primary operations and PTIR 95% lower confidence > 2 x group PTIR



## Probability of becoming a long term outlier



Outlier Level	Estimated probability of becoming a persistent outlier
Level 2+ (>1.5x group PTIR)	4.5%
Level 1 (>2x group PTIR)	0.6%

The probability of iTotal G2 XE and iTotal G2 (Bicondylar tray) becoming a persistent long-term outlier is estimated from all the NJR evidence to date, and will change as more evidence is accumulated. The probability will approach 0% or 100% in the long term with a large number of implants. The product is likely to be identified by the product outlier process if the probability calculated here exceeds 95%.



## **Reasons for Revision**

Reasons for revision of primary procedures in which the implant was used.

Reason for Revision	Revised <sup>†</sup>	Expected Revisions <sup>*</sup>	p value
Infection	1	1.75	1
Progressive Arthritis Remaining	0	0.28	1
Aseptic Loosening Femur	1	0.34	0.288
Aseptic Loosening Tibia	1	0.77	0.537
Aseptic Loosening Patella	1	0.19	0.172
Pain	0	0.96	1
Stiffness	0	0.60	1
Malalignment	0	0.50	1
Instability	2	0.92	0.234
Dislocation / Subluxation	0	0.19	1
Periprosthetic Fracture	0	0.14	1
Wear of Polyethylene Component	1	0.10	0.095
Lysis - Tibia	1	0.16	0.146
Lysis - Femur	1	0.10	0.093
Component Dissociation	0	0.05	1
Implant Fracture	0	0.02	1
Other / Not recorded	0	0.46	1
Total Revised	3	5.80	0.399

<sup>†</sup> multiple reasons may be listed for one revision procedure

Significantly better, p < 0.001Significantly better, p < 0.05Significantly worse p < 0.05Significantly worse p < 0.001



Components Revised	Number of procedures	iTotal G2 XE and iTotal G2 (Bicondylar tray)	All other TKR in NJR
Femoral and/or Tibial Revision	3	100.0%	71.4%
Isolated bearing exchange	0	0.0%	11.9%
Patella revision / resurfacing only	0	0.0%	15.6%
No component revision recorded	0	0.0%	1.2%



<sup>\*</sup> Adjusted for agegroup, gender and indications

#### **Revision Details**


NJR Index No	Procedure ID	Primary Date	Revision Date	Years implanted	Age at Primary	Revision Reasons	
1812172	1597990	09/11/2011	18/12/2015	4.11	48	aseptic loosening femur; aseptic loosening patella; aseptic loosenin	g tibia; Instability; LysisFemur; LysisTibia; WearOfPolyethyleneComponent
1892906	1677844	19/03/2014	15/10/2015	1.57	65	Instability	
2042935	1825319	11/11/2014	20/12/2014	0.11	74	infection	
	ļ						
	1						

Table contains Most recent 20 revisions



## **Patient - Time Incidence Rate**

#### Lead Surgeons implanting iTotal G2 XE and iTotal G2 (Bicondylar tray)




Each circle represents one surgeon. Red circles represent surgeons with a higher than expected revision rate (p < 0.001, and blue circles represent surgeons with a lower than expected revision rate (p < 0.001). Note that these rates have not been adjusted for case mix, or for variants of implant chosen.



## **Cumulative Revision Rate**

Endpoint: All reasons for revision



Cox Proportional Hazards model for revision risk ratio of iTotal G2 XE and iTotal G2 (Bicondylar tray) / All other TKR in NJR:

Adjustment	Hazard Ratio (95% CI)	p-value
Unadjusted	0.58 (0.19 - 1.81)	0.349
Adjusted for age, gender, year cohort and indications.	0.55 (0.18 - 1.71)	0.305



## **Cumulative Revision Rate**

**Endpoint: All revisions** 

Period /years	At Risk	iTotal G2 XE and iTotal G2 (Bicondylar tray)	All TKR in NJR
0	576	-	-
1	407	0.2% (0.0% - 0.6%)	0.4% (0.4% - 0.4%)
2	234	0.5% (0.1% - 1.3%)	1.0% (1.0% - 1.0%)
3	114	0.5% (0.1% - 1.7%)	1.5% (1.5% - 1.6%)
4	42	0.5% (0.1% - 2.4%)	1.9% (1.9% - 1.9%)
5		( - )	2.2% (2.1% - 2.2%)
6		( - )	2.4% (2.4% - 2.5%)
7		( - )	2.7% (2.6% - 2.7%)
8		( - )	2.9% (2.9% - 3.0%)
9		( - )	3.2% (3.1% - 3.2%)
10		( - )	3.4% (3.3% - 3.4%)

Cumulative revision rate with 95% confidence intervals Rate is only reported for times where more than 40 remain at risk



Summary Report for:

iTotal G2 XE and iTotal G2 (Bicondylar tray)

## Disclaimer

#### Disclaimer

This report is produced by Northgate Public Services (UK) Limited (NPS) using data collected, collated and provided by third parties. As a result, NPS takes no responsibility for the accuracy, currency, reliability and correctness of any data used or referred to in this report, and disclaims all warranties in relation to such data, in each case to the maximum extent permitted by legislation.

To the maximum extent permitted by law, NPS shall have no liability (including but not limited to liability by reason of negligence) for any loss, damage, cost or expense incurred or arising by reason of any person using or relying on the data within this report and whether caused by reason of any error, omission or misrepresentation in the report or otherwise. This report is not to be taken as advice. Third parties using or relying on the data in this report do so at their own risk and will be responsible for making their own assessment and should verify all relevant representations, statements and information with their own professional advisers.



# **Appendix A1 - Femoral Components**

#### 

M5725INT0600010 - FI	Left KNEE - Femoral Implant	142
M5725INT0600020 - FI	Right KNEE - Femoral Implant	171
M5725INT060XE10 - FI	LEFT KNEE - Femoral Implant	127
M5725INT060XE20 - FI	RIGHT KNEE - Femoral Implant	136



# **Appendix A2 - Tibial Components**

# M5725INT0600010 - TT Left KNEE - Tibial Tray 143 M5725INT0600020 - TT Right KNEE - Tibial Tray 169 M5725INT060XE10 - TT LEFT KNEE - Tibial Tray 127 M5725INT060XE20 - TT RIGHT KNEE - Tibial Tray 137



# **Appendix A3 - Bearing Components**

#### **Catalogue Number** Number Description M5725INT0600010 - 6MI Left KNEE - 6mm Medial Insert 27 M5725INT0600010 - 7MI Left KNEE - 7mm Medial Insert M5725INT0600010 - 8MI Left KNEE - 8mm Medial Insert M5725INT0600010 - ALI Left KNEE - A Lateral Insert 4 M5725INT0600010 - BLI Left KNEE - B Lateral Insert M5725INT0600020 - 6MI Right KNEE - 6mm Medial Insert 24 M5725INT0600020 - 7MI Right KNEE - 7mm Medial Insert 2 7 M5725INT0600020 - ALI Right KNEE - A Lateral Insert 5 M5725INT060XE10 - 6M LEFT KNEE - 6mm Medial Insert 6 M5725INT060XE10 - 8MILEFT KNEE - 8mm Medial Insert M5725INT060XE10 - ALI|LEFT KNEE - A Lateral Insert M5725INT060XE10 - BLILEFT KNEE - B Lateral Insert 4 M5725INT060XE10 - CLILEFT KNEE - C Lateral Insert 5 M5725INT060XE20 - 6M/RIGHT KNEE - 6mm Medial Insert 2 M5725INT060XE20 - 7M|RIGHT KNEE - 7mm Medial Insert M5725INT060XE20 - 8M RIGHT KNEE - 8mm Medial Insert 4 M5725INT060XE20 - ALI|RIGHT KNEE - A Lateral Insert 2 M5725INT060XE20 - BLIRIGHT KNEE - B Lateral Insert M5725INT060XE20 - CLIRIGHT KNEE - C Lateral Insert

